

| CATHETER PRIMING VOLUMES | | | | CATHETER FLUSH/DISCARD VOLUMES | | | |
|---------------------------------------|---------|-------------|-------------|---------------------------------------|---------|--------|-------|
| Brand | Size | small (vol) | large (vol) | Brand | Size | small | large |
| Tunneled Catheters | | | | Tunneled Catheters | | | |
| Bard Broviac (Single Lumen) | | | | Bard Broviac (Single Lumen) | | | |
| | 2.7F | 0.15ml | N/A* | | 2.7F | 0.45ml | N/A |
| | 4.2F | 0.3ml | N/A | | 4.2F | 0.9ml | N/A |
| | 6.6F | 0.7ml | N/A | | 6.6F | 2.1ml | N/A |
| Bard Hickman (Single Lumen) | | | | Bard Hickman (Single Lumen) | | | |
| | 9.6F | 1.8ml | N/A | | 9.6F | 5.4ml | N/A |
| BardHickman (Double Lumen) | | | | BardHickman (Double Lumen) | | | |
| | 7F | 0.6ml | 0.8ml | | 7F | 1.8ml | 2.4ml |
| | 9F | 0.6ml | 1.3ml | | 9F | 1.8ml | 3.9ml |
| | 10F | 1.3ml | 1.3ml | | 10F | 3.9ml | 3.9ml |
| | 12F | 1.8ml | 1.8ml | | 12F | 5.4ml | 5.4ml |
| Cook Catheter (Double Lumen) | | | | Cook Catheter (Double Lumen) | | | |
| | 7F | 0.7ml | 0.9ml | | 7F | 2.1ml | 2.7ml |
| | 9F | 1.5ml | 2.0ml | | 9F | 4.5ml | 6.0ml |
| | 12F | 1.6ml | 2.7ml | | 12F | 4.8ml | 8.1ml |
| Non-tunneled Catheters | | | | Non-tunneled Catheters | | | |
| Cook Catheters (Single lumen) | | | | Cook Catheters (Single lumen) | | | |
| | 2.5F | 0.05ml | N/A | | 2.5F | 0.15ml | N/A |
| | 3F | 0.1ml | N/A | | 3F | 0.3ml | N/A |
| | 4F | 0.1ml | N/A | | 4F | 0.3ml | N/A |
| Cook Catheter (Double lumen) | | | | Cook Catheter (Double lumen) | | | |
| | 4F | 0.1ml | 0.2ml | | 4F | 0.3ml | 0.6ml |
| | 5F | 0.2ml | 0.2ml | | 5F | 0.6ml | 0.6ml |
| Cook Catheter (Triple lumen) | | | | Cook Catheter (Triple lumen) | | | |
| | 5F | 0.2ml | 0.3ml | | 5F | 0.6ml | 0.9ml |
| | 7F | 0.3ml | 0.5ml | | 7F | 0.9ml | 1.5ml |
| Implanted Venous Access Lumens | | | | Implanted Venous Access Lumens | | | |
| Bard Single Lumens | | | | Bard Single Lumens | | | |
| Low profile | 6.6F | 0.9ml | N/A | Low profile | 6.6F | 2.7ml | N/A |
| X-Port ISP | 6Fr | 1.6ml | N/A | X-Port ISP | 6Fr | 4.8ml | N/A |
| Regular | 6.6F | 1.2ml | N/A | Regular | 6.6F | 3.6ml | N/A |
| Regular | 8F | 1.8ml | N/A | Regular | 8F | 5.4ml | N/A |
| Bard Dual Lumens | | | | Bard Dual Lumens | | | |
| | 10F MRI | 1.8ml | 1.8ml | | 10F MRI | 5.4ml | 5.4ml |

| Peripherally Inserted Central Catheters (PICCs) | | | | Peripherally Inserted Central Catheters (PICCs) | | | |
|--|--------|--------|--|---|--------|--------|--|
| Arrow (Dual lumen) | | | | Arrow (Dual lumen) | | | |
| 4F | 0.33ml | 0.38ml | | 4F | 0.99ml | 1.14ml | |
| Bard Groshong (Single lumen) | | | | Bard Groshong (Single lumen) | | | |
| 3F | 0.31ml | N/A | | 3F | 0.93ml | N/A | |
| 4F | 0.49ml | N/A | | 4F | 1.47ml | N/A | |
| Bard Groshong (Double lumen) | | | | Bard Groshong (Double lumen) | | | |
| 5F | 0.41ml | 0.41ml | | 5F | 1.23ml | 1.23ml | |
| BD L-cath | | | | BD L-cath | | | |
| 1.2F (28G) | 0.06ml | N/A | | 1.2F (28G) | 0.18ml | N/A | |
| 1.9F (24G) | 0.1ml | N/A | | 1.9F (24G) | 0.3ml | N/A | |
| HDC V-cath (Single lumen) | | | | HDC V-cath (Single lumen) | | | |
| 2F | 0.08ml | N/A | | 2F | 0.24ml | N/A | |
| 3F | 0.22ml | N/A | | 3F | 0.66ml | N/A | |
| 4F | 0.38ml | N/A | | 4F | 1.14ml | N/A | |
| Boston Sci. Vaxcel (PASV) (Single lumen) | | | | Boston Sci. Vaxcel (PASV) (Single lumen) | | | |
| 4F | 1.5ml | N/A | | 4F | 4.5ml | N/A | |
| Boston Sci. Vaxcel (PASV) (Double lumen) | | | | Boston Sci. Vaxcel (PASV) (Double lumen) | | | |
| 5F | 1ml | N/A | | 5F | 3ml | 3ml | |
| Bard Per-Q-cath (Single lumen) | | | | Bard Per-Q-cath (Single lumen) | | | |
| 3F | 0.38ml | N/A | | 3F | 1.14ml | N/A | |
| Bard Per-Q-cath (Double lumen) | | | | Bard Per-Q-cath (Double lumen) | | | |
| 4F | 0.38ml | 0.45ml | | 4F | 1.14ml | 1.35ml | |
| Angiodynamics Morpheus CT PICC | | | | Angiodynamics Morpheus CT PICC | | | |
| 3F | 0.34ml | N/A | | 3F | 1.02ml | N/A | |
| 4F | 0.52ml | N/A | | 4F | 1.56ml | N/A | |
| 5F | 0.82ml | N/A | | 5F | 2.46ml | N/A | |
| See product package/insert for priming volumes of add-on devices | | | | | | | |
| * N/A=Not Applicable | | | | | | | |
| | | | | | | | |